

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 7 of 12

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 EPA Lab Code: POB 25046-MS 407-BLDG 95
 DENVER, CO-80225-0046
 CO00009 UNITED STATES
 (303) 236-2000
 Director: MR. MERLE SHOCKEY

Shipment Date : 18-Jul-2005
 Closing Date : 01-Sep-2005
 Score Date : 21-Sep-2005

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
 NVLAP Lab Code 200387-0.

Shipment: 285 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Base/Neutrals								
Bis(2-ethylhexyl)phthalate EPA Code: N/A	ug/L	8536	USGS O-3116-87	50.1	96.3		34.2 - 158	Satisfactory 73 passed out of 75 reported results.
Benzo(a)anthracene EPA Code: N/A	ug/L	8536	USGS O-3116-87	57.4	65.2		34.3 - 96.1	Satisfactory 78 passed out of 79 reported results.
Benzo(a)pyrene EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 1.3	< 20.0 *			Satisfactory 79 passed out of 80 reported results.
Benzo(b)fluoranthene EPA Code: N/A	ug/L	8536	USGS O-3116-87	36.3	40.8		16.9 - 64.6	Satisfactory 79 passed out of 80 reported results.
Benzo(g,h,i)perylene EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 1.6	< 20.0 *			Satisfactory 77 passed out of 79 reported results.
Benzo(k)fluoranthene EPA Code: N/A	ug/L	8536	USGS O-3116-87	58.5	73.0		21.4 - 125	Satisfactory 80 passed out of 80 reported results.
Benzyl butyl phthalate EPA Code: N/A	ug/L	8536	USGS O-3116-87	73.0	69.6		14.0 - 125	Satisfactory 67 passed out of 74 reported results.
Chrysene EPA Code: N/A	ug/L	8536	USGS O-3116-87	85.8	86.0		40.1 - 132	Satisfactory 78 passed out of 79 reported results.
Dibenzo(a,h)anthracene EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 2.2	< 20.0 *			Satisfactory 77 passed out of 78 reported results.
Di-n-butyl phthalate EPA Code: N/A	ug/L	8536	USGS O-3116-87	134	133		53.6 - 213	Satisfactory 71 passed out of 74 reported results.
Diethyl phthalate EPA Code: N/A	ug/L	8536	USGS O-3116-87	110	104		25.1 - 182	Satisfactory 69 passed out of 74 reported results.
Dimethyl phthalate EPA Code: N/A	ug/L	8536	USGS O-3116-87	< 1.0	< 100 *			Satisfactory 73 passed out of 74 reported results.